



## **MEDOS HILITE<sup>®</sup>**


*Reservoirs for Infants and Paediatrics*



# MEDOS HILITE リザーバーシリーズ

## 開発コンセプト

- スムーズな血流
- 視認性の向上
- より使いやすいポート設計



### Hardshell Venous Reservoir for Infants Paediatrics

	MVC 0730	MVC 1630
■ Operating volume	50mL - 700mL	50mL - 1600mL
■ Blood flow rate		
Venous section	1.2L/min	3L/min
Cardiotomy section	0.9L/min	2L/min
■ Cardiomy filtration	30 μm	30 μm
■ Port configurations		
Venous outlet	1x3/16"-1/4"	1x1/4"
<静脈脱血部ポート>		
Venous inlet	1x1/4"	1x3/8"
Luer Lock	2x female	2x female
<カーディオトミーポート>		
Sucker	4x3/16"-1/4", 2x1/4"	4x3/16"-1/4", 2x1/4"
Quick prime	1x1/4"	1x1/4"
Luer Lock	4x female	4x female
<フィルター不通過ポート>		
Vent	1x1/4"	1x1/4"
Luer Lock	2x female	2x female
■ Temperature port	at the venous inlet port	at the venous inlet port
■ アクセサリー	サンプリングライン	サンプリングライン、アダプター3/8-1/4"



### Cardiotomy - Reservoir for Infants Paediatrics

	MC 0740	MC 1640
■ Operating volume	50mL - 700mL	50mL - 1600mL
■ Blood flow rate	0.9L/min	2L/min
■ Cardiomy filtration	40 μm	40 μm
■ Port configurations		
Venous outlet	1x3/16"-1/4"	1x1/4"
<フィルター通過ポート>		
Sucker	4x3/16"-1/4", 2x1/4"	4x3/16"-1/4", 2x1/4"
Quick prime	1x1/4"	1x1/4"
Luer Lock	4x female	4x female
<フィルター不通過ポート>		
Vent	1x1/4"	1x1/4"
Luer Lock	2x female	2x female

製品名	製品番号	JANコード
MEDOS リザーバー MVC1630	ME 6221-0210	4571102333410
MEDOS リザーバー MVC0730	ME 6221-0310	4571102333427
MEDOS リザーバー MC 1640	ME 6222-0212	4571102333380
MEDOS リザーバー MC 0740	ME 6222-0312	4571102333403

医療機器認証番号 21600BZY00618000

■ 製造販売業者

**株式会社トライテック** (<http://www.trytech.co.jp>)

〒140-0013 東京都品川区南大井1-20-17 マルイト南大井ビル3F  
 営業本部 / TEL 03-5767-9831 FAX 03-5767-9839  
 関 西 / TEL06-6535-7163 FAX06-6535-7164

■ 製造業者: MEDOS medizintechnik AG (ドイツ)

(株)トライテックは ISO 13485 の認証を取得し、このシステムに基づき製品管理への取り組みを着実に進めております。

TRT-MED-HR P-1511-0001

